

16103057

9.

510(K) SUMMARY

9.0 510(K) SUMMARY

NOV 19 2010

This 510(k) summary of safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.92.

**APPLICANT** Asahi Intecc Co., Ltd.  
1703 Wakita-cho, Moriyama-ku  
Nagoya, Aichi 463-0024  
Japan

**OFFICIAL CORRESPONDENT** Yoshi Terai  
President, CEO  
Asahi Intecc USA, Inc.  
2500 Red Hill Avenue, Suite 210  
Santa Ana, CA 92705  
Tel: (949) 756-8252  
FAX (949) 756-8165  
e-mail: yoshi.terai@asahi-intecc.com

**TRADE NAME:** ASAHI Astato XS 20 Peripheral Guide Wire

**COMMON NAME:** Guide Wire

**CLASSIFICATION NAME:** Wire, Guide, Catheter

**DEVICE CLASSIFICATION:** Class 2 per 21 CFR §870.1330

**PRODUCT CODE** DQX

**PREDICATE DEVICE:** Asahi – Astato 30 Peripheral Guide Wire – 510(k) K071721  
Asahi – Treasure Guide Wire – 510(k) K061984  
Asahi – JoWire Neo's PTCA Guide Wire – 510(k) K022762  
Asahi – Asahi PTCA Guide Wire (Miracle bros 12/Confianza Pro 12) – 510(k) K052339  
Asahi – Asahi PTCA Guide Wire (PTFE coating type II) – 510(k) K070945

**DESCRIPTION OF THE DEVICE SUBJECT TO PREMARKET NOTIFICATION:**

The ASAHI Astato XS 20 Peripheral Guide Wire is a steerable guide wire with a maximum diameter of 0.014 inches (0.36mm) and available in 180 cm and 300 cm length. The guide wire is constructed from a stainless steel core wire with platinum-nickel coil. The coil part (distal end) of the guide wire is radiopaque to achieve visibility. The distal end of the coil part is a straight configuration and is easily bendable with the vessel curvature.

A hydrophilic coating is applied to the coil part (distal portion) of the guide wire. The proximal section of the guide wire is coated with PTFE, and the PTFE in the proximal coating is available in two types - PTFE type I and type II.

CONFIDENTIAL  
Asahi Intecc  
October 14, 2010

**INDICATION FOR USE:**

The Asahi Astato XS 20 Peripheral Guide Wire is intended to facilitate the placement and exchange of diagnostic and therapeutic devices during intravascular procedures. This device is intended for peripheral vascular use only.

---

**TECHNICAL CHARACTERISTICS:**

This submission represents minor dimensional specifications and the addition of a PTFE coating. All other specifications are equivalent to those listed for the currently cleared predicate devices.

Comparisons of the ASAHI Astato XS 20 Peripheral Guide Wire and predicate devices show that the technological characteristics such as product performance, design and intended use are substantially equivalent to the current marketed predicate devices.

The ASAHI Astato XS 20 Peripheral Guide Wire is similar in design - device dimensional specifications, and intended use, manufacturing process, operating principle, shelf life and sterilization process are the same and materials that have been used in other predicate devices in that its core wire, tip coils and solders remain the same.

---

**PERFORMANCE DATA:**

Enclosed within this submission is performance data that demonstrates that the Asahi Astato XS 20 Peripheral Guide Wire meets all predetermined performance criteria.

All components that come in direct contact with the patient have a long history of use in medical devices and are proven to be biocompatible for use in the vasculature. Furthermore, this submission contains reference to predicate ASAHI devices that use the same materials as used in the subject device.

And In vitro bench testing and shelf-life testing, including tensile strength, torque strength, torqueability, tip flexibility, coating adherence and catheter compatibility as listed below were conducted on the ASAHI Astato XS 20 Peripheral Guide Wire. This 510(k) notice includes mechanical and functional bench testing that demonstrates that the ASAHI Astato XS 20 Peripheral Guide Wire performs as intended.

Performance test/evaluation summary:

Device performance:

Tensile Strength

Turns to Failure (Torque Strength)

Torqueability (Torque Response)

Tip Flexibility

Coating Adhesion

Slipping Ability of Guide Wire in PTA Balloon Catheter

Particulate testing

Biocompatibility:

Systemic Toxicity Study

In Vitro Hemolysis Study

---

CONFIDENTIAL  
Asahi Intecc  
October 14, 2010

Intracutaneous Study  
Cytotoxicity Study  
Sensitization Study  
Pyrogen Study  
Plasma Recalcification Time Coagulation Study  
In Vivo Thromboresistance Study  
C3a Complement Activation Study  
SC5b-9 Complement Activation Study

---

**SUMMARY/CONCLUSION:**

The Asahi Astato XS 20 Peripheral Guide Wire characteristics are substantially equivalent to the specified predicate device and other currently marketed devices for the same indication for use.

---



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Room --WO66-G609  
Silver Spring, MD 20993-0002

Asahi Intecc Co., Ltd.  
c/o Mr. Yoshi Terai  
President, CEO  
2500 Red Hill Avenue, Suite 210  
Santa Ana, CA 92705

NOV 19 2010

Re: K103057

Trade/Device Name: Astatto XS 20 Peripheral Guide Wire  
Regulation Number: 21 CFR 870.1330  
Regulation Name: Catheter guide wire  
Regulatory Class: Class II (two)  
Product Code: DQX

Dated: October 14, 2010

Received: October 15, 2010

Dear Mr. Terai:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.


If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOffices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



 Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

2.

**INDICATIONS FOR USE STATEMENT**

2.0 INDICATIONS FOR USE STATEMENT

510(k) Number (if known): K103057

NOV 19 2010

Device Name: ASAHI Astato XS 20 Peripheral Guide

Indications for Use:

The ASAHI Astato XS 20 Peripheral Guide Wire is intended to facilitate the placement and exchange of diagnostic and therapeutic devices during intravascular procedures. This device is intended for peripheral vascular use only.

Prescription Use X  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 801 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE OF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Dwight R. Johnson  
(Division Sign-Off)  
Division of Cardiovascular Devices

Page 1 of 1

510(k) Number K103057